



Ref. Certif. No.

DK-139892-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product Current transducers for contact less measurement of AC/DC currents

Name and address of the applicant
Danisense A/S
Malervej 10 Taastrup 2630
Denmark

Name and address of the manufacturer
Danisense A/S
Malervej 10 Taastrup 2630
Denmark

Name and address of the factory
Danisense A/S
Malervej 10 Taastrup 2630
Denmark

Note: When more than one factory, please report on page 2

Ratings and principal characteristics
DSSIU-6-1U: 100-240V~, 50-60 Hz, Max. 350W.
☒ [Additional Information on page 2](#)

Trademark / Brand (if any)
DANISENSE

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.
DT100ID, DN1000ID, DM1200UB-10V, DL2000UB-10V, DS200UB-10V
☒ [Additional Information on page 2](#)

Additional information (if necessary may also be reported on page 2)
Additionally evaluated to: EN 61010-1:2010, EN 61010-1:2010/A1:2019
National Differences specified in the CB Test Report.

A sample of the product was tested and found to be in conformity with
IEC 61010-1:2010, IEC 61010-1:2010/AMD1:2016, IEC 61010-2-030:2017

As shown in the Test Report Ref. No. which forms part of this Certificate
E502957-D1000-2/A0/C0-CB issued on 2023-04-13

This CB Test Certificate is issued by the National Certification Body



- ☐ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☒ UL Solutions (Denmark), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2023-04-14

Signature:
Jan-Erik Storgaard



Ref. Certif. No.

DK-139892-UL

Additional Model Detail(s):

Interface unit.: Models: DSSIU-6-1U

Current Transducers: DS50UB-10V, DS200UB-10V, DS600UB-10V, DL2000UB-10V, DT100ID, DN1000ID, DM1200UB-10V

Additional Ratings:

The current transducers are designed for operation with a $\pm 15V$ DC power supplies provided by the DSSIU-6-1U. Voltage rating are not required on the current transducers. The product datasheets describes the correct voltage to connect.

Additional information (if necessary)



- ☐ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☒ UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2023-04-14

Signature:

Jan-Erik Storgaard